

Appendix C

Calibration Logs

Work Order No.: SE-062487
Date of Service: 11/27/2018 12:00:00 AM

Unit Under Test: YSI ProDSS, 4M with Flow Cell

Asset No.: FA03151

David Perkinson

Technician Initials: DP

Serial No: 17C105115/17D105931

TEST	Specification	Result
Standard Calibration	Pass/Fail	PASS

TEST STANDARDS USED:

DESCRIPTION	LOT NO./EXPIRATION DATE	QUANTITY
pH 10.00 Standard Solution	Lot No. C800938 Exp. 10/05/2019	1
pH 4.00 Standard Solution	Lot No. 800499 Exp. 9/19/2019	1
pH 7.00 Standard Solution	Lot No. 18262594 Exp. 05/20	1
7.00 mS Conductivity Standard Solution	Lot: 8GJ311 Exp: OCT 2019	1
Turbidity Free Water		1
124 FNU YSI Turbidity Standard	Lot: 18H18303752 exp. 08/2019	1
ORP Standard Solution	Lot No. 17K100603 Exp. 01/19/2019	1
Sodium Sulfitite/ Zero DO Standard	Lot No. C473638	1

TEST EQUIPMENT USED:

DESCRIPTION	ASSET NO.	SERIAL NO.	DATE OF LAST CAL	DATE CAL DUE

Test Equipment and standards are traceable to National standards.



Calibration Certificate

rev 8/9/11

Work Order No.: SE-063862

Date of Service: 1/14/2019 12:00:00 AM

Unit Under Test: YSI ProDSS, 4M with Flow Cell

Asset No.: FA03699

David Perkinson

Technician Initials:

Serial No: 18H110750/18H108461

TEST	Specification	Result
Standard Calibration	Pass/Fail	PASS

TEST STANDARDS USED:

DESCRIPTION	LOT NO./EXPIRATION DATE	QUANTITY
pH 10.00 Standard Solution	Lot No. C800938 Exp. 10/05/2019	1
pH 4.00 Standard Solution	Lot No. C62576 Exp.9/18/2020	1
pH 7.00 Standard Solution	Lot No. 18262594 Exp. 05/20	1
7.00 mS Conductivity Standard Solution	Lot: 8GK926 Exp: Nov 2019	1
ORP Standard Solution	Lot No. 17K100603 Exp. 01/19/2019	1
Sodium Sulfite/ Zero DO Standard	Lot No. C473638	1
Turbidity Free Water		1
124 FNU YSI Turbidity Standard	Lot: 18H18303752 exp. 08/2019	1
Air Saturated Water		1

TEST EQUIPMENT USED:

DESCRIPTION	ASSET NO.	SERIAL NO.	DATE OF LAST CAL	DATE CAL DUE

Test Equipment and standards are traceable to National standards.